

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 4490 (1968): Forceps, Dilating, Trousseau's [MHD 1: Surgical Instruments]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS : 4490 - 1968

Indian Standard

REAFFIRMED 7/8/86

**SPECIFICATION FOR
FORCEPS, DILATING, TROUSSEAU'S**

UDC 615.472.41



© Copyright 1968

**INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 1**

Revised Price **Gr 2**

March 1968

Indian Standard

SPECIFICATION FOR FORCEPS, DILATING, TROUSSEAU'S

Surgical Instruments Sectional Committee, CPDC 11

Chairman

COL R. D. AYYAR

Representing

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

*Members*CHIEF SALES MANAGER
WORKS MANAGER (*Alternate*)

SHRI R. D. GUPTA

DR (KUMARI) S. D. GUPTA

BRIG IQBAL CHAND

SHRI G. B. JAKHETIA

SHRI V. KRISHNAMOORTHY

SHRI SANTI KUMAR MUKERJI

SHRI M. G. PANDIT

DR A. RAMDAS

SHRI S. SEN GUPTA

SHRI MAHAJAN M. SHAH

MAJ SOHAN SINGH

SHRI K. L. TALWAR

SHRI A. N. VIG

SHRI R. N. VIG (*Alternate*)

SHRI A. B. RAO,

Director (Consr Prod) (*Secretary*)

Indian Drugs & Pharmaceuticals Ltd, New Delhi

National Metallurgical Laboratory (CSIR),
JamshedpurLady Hardinge Medical College & Hospital,
New DelhiDirectorate General of Armed Forces Medical
Services (Ministry of Defence)

Development Commissioner, Small Scale Industries

Directorate General of Technical Development

H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta

Ministry of Health & Family Planning

Ministry of Railways

Baruipur Surgical Instrument Centre (Directorate of
Industries, Government of West Bengal)

Shah Medical & Surgical Co Ltd, Baroda

Directorate General of Inspection (Ministry of
Defence)Chief Inspectorate of General Stores [Ministry of
Defence (DGI)], Kanpur

ISCOMED Industries Private Ltd, Bombay

Director General, ISI (*Ex-officio Member*)

(Continued on page 2)

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 1

IS : 4490 - 1968

(Continued from page 1)

Precision Type Surgical Instruments Subcommittee, CPDC 11 : 5

Convener

DR J. L. GUPTA

Representing

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

Members

DR JACOB CHANDI

Christian Medical College, Vellore

DR G. S. GOPALAN

Jaya Surgicals, Madras

SHRI G. JAYA PRASAD (Alternate)

DR GOPINATH

All India Institute of Medical Sciences, New Delhi

BRIG IQBAL CHAND

Directorate General of Armed Forces Medical Services (Ministry of Defence)

SHRI SANTI KUMAR MUKERJI

H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta

SHRI H. C. MUKHERJEA

Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

SHRI J. PRABHAKAR

Indian Drugs & Pharmaceuticals Ltd, Madras

DR T. T. RAMALINGAM

In personal capacity (4-A Kilpauk Garden Road, Kilpauk, Madras)

DR P. B. RAO

Jawaharlal Institute of Post-Graduate Medical Education & Research (Ministry of Health & Family Planning), Pondicherry

Indian Standard

SPECIFICATION FOR FORCEPS, DILATING, TROUSSEAU'S

0. FOREWORD

0.1 This Indian Standard was adopted by the Indian Standards Institution on 29 January 1968, after the draft finalized by the Surgical Instruments Sectional Committee had been approved by the Consumer Products Division Council.

0.2 The formulation of Indian Standards on surgical instruments has been taken up on the recommendations of the Advisory Committee for the Development of Surgical Instruments, Equipment and Appliances, Government of India.

0.3 This Indian Standard is one of a series of Indian Standards on surgical instruments. Other standards published so far in the series appear at P 6.

0.4 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS : 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard lays down the requirements for Trousseau's dilating forceps used in ENT work.

2. MATERIAL

2.1 The forceps and spring shall be made from stainless steel conforming to Designation 30Cr13 of Schedule V of IS : 1570-1961†.

2.2 The screws shall be of stainless steel conforming to the Designation mentioned in 2.1 or 20Cr18Ni2 of Schedule V of IS : 1570-1961†.


3. SHAPE AND DIMENSIONS

3.1 The general shape and dimensions of the forceps shall be as indicated in Fig. 1.

*Rule for rounding off numerical values (*revised*).

†Schedules for wrought steels for general engineering purposes.

3.2 Joint — The joint of the forceps shall be as given in Fig. 1. The other requirements for the joint shall be the same as in 6 of Section 2 of IS : 3642-1966*.



NOTE — For finger loop, see IS : 3642-1966 General requirements for surgical instruments.

(All dimensions in millimetres)

FIG. 1 DILATING FORCEPS, TROUSSEAU'S PATTERN

*General requirements for surgical instruments.

4. WORKMANSHIP AND FINISH

4.1 The forceps shall be of a symmetrical and balanced construction, and shall open and close without jerk. It shall be capable of opening fully when a force of 250 to 400 gf is applied on one finger loop approximately at its equator, while the other arm is fixed. The jaws shall close accurately. The forceps shall be free from surface defects like scales, burrs, pits, cracks and seams. All sharp edges shall be rounded smooth. The forceps shall be passivated and polished bright.

5. HEAT-TREATMENT

5.1 The forceps and its spring shall be suitably hardened and tempered to give a uniform hardness of 430 to 490 HV.

6. TESTS

6.1 Flexibility — The finger loops of the forceps shall be pressed completely in a suitable device. The tips shall then be pressed so as to bring them closer by 5 mm measured at the extreme end and released. The forceps shall not have any permanent set after the test.

6.2 Corrosion Resistance — The forceps shall be tested for corrosion resistance as given in 6.2.1 and 6.2.2.

6.2.1 Copper Sulphate Test — The sample shall be scrubbed with soap and warm water, rinsed in hot water, followed by a dip in ethyl alcohol (95 percent) and dried. The sample shall be completely immersed in copper sulphate solution at room temperature for six minutes and shall then be washed off with fresh water or wet cotton wool. The copper sulphate solution shall be made up as follows:

Copper sulphate ($\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$)	4.0 g
Sulphuric acid (H_2SO_4) (sp gr 1.84)	10.0 g
Water (H_2O)	90.0 ml

6.2.2 There shall be no red stains or spots on the sample after the test, but the polished surface getting dull may be permitted.

6.3 Spring — The spring shall be tested by pressing either arm to the extreme position and released. On repeating this for 50 times, there shall be no permanent set observed and the jaws shall not gape.

7. MARKING

7.1 Each forceps shall be clearly and indelibly marked with the manufacturer's name, initials or trade-mark, in addition to the words 'STAINLESS STEEL'.

7.1.1 The forceps may also be marked with the ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products as actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

8. PACKING

8.1 The forceps shall be wrapped in moisture-proof paper or polyethylene bag; then they shall be packed in a manner that coming into contact with each other is avoided.

INDIAN STANDARDS ON Surgical Instruments

IS :		Rs
3318-1965	General requirements for surgical scalpels and knives ...	1-00
3319-1965	Surgical scalpels with detachable blades (Bard Parker type) ...	3-00
3320-1965	Surgical scalpels ...	1-00
3642-1966	General requirements for surgical instruments ...	2-50
3643-1966	Dissecting forceps (serrated and toothed) ...	2-00
3644-1966	Artery forceps (Halsted's mosquito pattern) ...	1-50
3645-1966	Artery forceps (Spencer Wells pattern) ...	1-50
3758-1966	Hooks, aural ...	1-50
3759-1966	Blades, skin grafting ...	1-00
3760-1966	Probe, aural, Jobson Harne's pattern ...	1-00
3767-1966	Hook, skin, Gillies pattern ...	1-00
3787-1966	Gag, mouth, cleft palate ...	1-50
3788-1966	Specula, aural ...	1-50
3799-1966	Forceps, aural and nasal, crocodile action ...	2-00
3807-1966	Forceps, dressing, aural and nasal ...	1-50
4085-1967	Knives, amputating ...	2-00
4089-1967	Forceps, clip removing Michel's ...	2-50
4094-1967	Forceps, sterilizer (Cheate's) ...	2-50
4244-1967	Retractors, skin, double ended ...	2-00
4245-1967	Needle holder (Gillies) combined with scissors, for plastic surgery ...	2-50
4275-1967	Scissors, Kilner's for plastic surgery ...	2-50
4281-1967	Scissors, Mc Inode's, for plastic surgery ...	3-00
4282-1967	Forceps, dissecting, plastic surgery ...	3-50
4487-1968	Forceps, tonsil artery (Birkett's pattern) (<i>under print</i>) ...	—
4488-1968	Forceps, tonsil holding (Muck's and White's pattern) (<i>under print</i>) ...	—
4489-1968	Forceps, nasal turbinate, Luc's (<i>under print</i>) ...	—
4490-1968	Forceps, dilating, Troussieu's ...	2-00

PUBLICATIONS OF INDIAN STANDARDS INSTITUTION

INDIAN STANDARDS

About 4 500 Indian Standards, broadly classified under the following main heads, have been issued so far:

Agriculture & Food
Chemical
Civil Engineering
Consumer Products

Electrotechnical
Mechanical Engineering
Structural & Metals
Textile

Of these, the standards belonging to the Consumer Products Group, at present, fall under the following categories:

Coir and Coir Products
Domestic Hardware
**Medical Instruments and
Hospital Equipment**
**Oil and Gas Burning
Appliances**

Safety Matches
**Sports Goods and Gymnasium
Equipment**
Umbrellas
Utensils and Cutlery
Writing Aids

OTHER PUBLICATIONS

	Ra
ISI Bulletin (Published Every Month)	
Single Copy	2.00
Annual Subscription	15.00
Annual Reports (from 1948-49 Onwards)	2.00 to 3.00 each
Handbook of ISI Publications, 1967	5.00

Available from

INDIAN STANDARDS INSTITUTION

Headquarters

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 1

Telephones 27 36 11 - 20

Telegrams Manaksanstha

Branch Offices

Telegrams Manaksanstha

534 Sardar Vallabhbhai Patel Road
5 Chowringhee Approach
Industrial Estate Administrative
Building, Sanatnagar
117/418 B Sarvodaya Nagar
54 General Patters Road

Bombay 7	Telephone	35 69 44
Calcutta 13	"	23-18 23
Hyderabad 18	"	3 92 29
Kanpur	"	3 76 95
Madras 2	"	8 72 78

AMENDMENT NO. 1 JULY 1975
TO
IS:4490-1968 SPECIFICATION FOR FORCEPS,
DILATING, TROUSSEAU'S

Alterations

(Page 5, clauses 6.2 and 6.2.1) - Substitute the following for the existing clauses:

'6.2 Corrosion Resistance - The instrument shall satisfy the boiling and autoclaving test as mentioned in IS:7531-1975*.'

(Page 5, foot-note) - Insert the following foot-note at the bottom of the page:

'*Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments.'

(CPDC 11)

Reprography Unit, ISI, New Delhi